



PORTABLE APPLIANCE TESTERS

PAT-820/815/810

NEW!



PAT-8XX digital meters series are used to measure the parameters of portable electrical equipment (power tools, white goods, etc.) which determine their safety: **resistance of protective conductors, insulation resistance, continuity of connections, leakage current, RCD devices, power.** Also PAT-820 meter allows flash test / high voltage test.

Meters can be used to test the equipment performed in accordance with standards:

- EN 60745-1 Hand-held motor-operated electric tools. Safety. General requirements
- EN 61029 Safety of transportable motor-operated electric tools - General requirements
- EN 60335-1 Household and similar electrical appliances -Safety -Part 1: General requirements
- EN 60950 Safety of information technology equipment (IT Equipment)
- AS/NZS 3760:2010 In-service safety inspection and testing of electrical equipment
- VDE 0404-1 Prüf - und Messeinrichtungen zum Prüfen der elektrischen Sicherheit von elektrischen Geräten. Teil 1: Allgemeine Anforderungen
- VDE 0404-2 Prüf - und Messeinrichtungen zum Prüfen der elektrischen Sicherheit von elektrischen Geräten. Teil 2: Prüfeinrichtungen für Prüfungen nach Instandsetzung, Änderung oder für Wiederholungsprüfungen
- VDE 0701-0702 Prüfung nach Instandsetzung, Änderung elektrischer Geräte. Wiederholungsprüfung elektrischer Geräte. Allgemeine Anforderungen für die elektrische Sicherheit

Functionality of the meter:

- intuitive user interface,
- manual tests and auto tests, the ability to describe auto test with standards or any name,
- flash test / high voltage test (only PAT-820),
- typing with QWERTY keyboard on the touch screen,
- description of test equipment, measurement location, customer data, assigning the serial number of the device under test and the index can be stored in meter memory, ability to write notes about the device under the test
- base of the appliances, customers, description of the equipment and damage.
- the results can be printed (also automatically after every measurement), reports (works with the printer), two labels can be printed after a single test (for the device and a removable wire)
- support for barcode reader (including 2D), readings of the original serial numbers for the appliances and registration codes and auto test codes,
- automatic measurement of RCD parameters
- build-in help with instructions how to connect test equipment and how to perform measurement,
- ability to create many user accounts with log-in function (as an option),
- supports USB flash drive;
- communication with PC via USB and WiFi;
- works with the program Sonel PAT Reader and Sonel PAT; measurement and settings configuration from the meter and also from PC, data analysis.

Sonel S.A.
 Wokulskiego 11
 58-100 Świdnica, PL
 tel. +48 74 85 83 860
 fax +48 74 85 83 809

export@sonel.pl
www.sonel.pl

Basic functions:

- measurement of earth bond/continuity resistance with the currents: 200 mA (PAT-810/815/820), 10 A and 25 A (only PAT-815/820) (protection class I),
- measurement of insulation resistance – three measurement voltages: 100 V and 250 V (only PAT-815/820), 500 V (PAT-810/815/820),
- measurement of substitute leakage current,
- measurement of differential leakage current,
- measurement of touch leakage current,
- measurement of power,
- measurement of current consumption,
- IEC lead test,
- check of the L-N circuit test,
- measurement of mains voltage and frequency,
- RCD testing,
- flash test/high voltage test (only PAT-820).

Other:

- automatic measurement range selection,
- professional software for data processing and reporting,
- cooperation with a barcode reader and printer,
- supports pendrive flash memory,
- large and clear touch display,
- ergonomic operation.

**Measurement of continuity resistance $I=200 \text{ mA}$
(protection class I)**

Display range	Resolution	Accuracy
0,00...0,99 Ω	0,01 Ω	$\pm(4\% \text{ m.v.} + 2 \text{ digits})$
1,00...19,99 Ω		$\pm(4\% \text{ m.v.} + 3 \text{ digits})$

- test current: $I=200 \text{ mA}$ for $R = 0,2...1,99 \Omega$
- adjustable limit
- adjustable measurement time

**Measurement of earth bond $I=10 \text{ A}$
(protection class I)**

Display range	Resolution	Accuracy
0...999 m Ω	1 m Ω	$\pm(3\% \text{ m.v.} + 4 \text{ digits})$
1,00...1,99 Ω	0,01 Ω	

- technical method
- test current: $I=10 \text{ A}$ for $R \leq 0,5 \Omega$
- adjustable limit
- adjustable measurement time

**Measurement of earth bond $I=25 \text{ A}$
(protection class I)**

Display range	Resolution	Accuracy
0...999 m Ω	1 m Ω	$\pm(3\% \text{ m.v.} + 4 \text{ digits})$
1,00...1,99 Ω	0,01 Ω	

- technical method
- test current: $I=25 \text{ A}$ for $R \leq 0,2 \Omega$
- adjustable limit
- adjustable measurement time

Measurement of insulation resistance

Measurement range according to IEC 61557-2 for:

$U_n=100 \text{ V}$: $100 \text{ k}\Omega...99,9 \text{ M}\Omega$

$U_n=250 \text{ V}$: $250 \text{ k}\Omega...199,9 \text{ M}\Omega$

$U_n=500 \text{ V}$: $500 \text{ k}\Omega...599,9 \text{ M}\Omega$

U_n	Range	Resolution	Accuracy
100V	0...1999k Ω	1k Ω	$\pm(5\% \text{ m.v.} + 8 \text{ digits})$
	2,0...19,99M Ω	0,01M Ω	
	20,0...99,9M Ω	0,1M Ω	
250V	0...1999k Ω	1k Ω	$\pm(5\% \text{ m.v.} + 8 \text{ digits})$
	2,00...19,99M Ω	0,01M Ω	
	20,0...99,9M Ω	0,1M Ω	
500V	0...1999k Ω	1k Ω	$\pm(5\% \text{ m.v.} + 8 \text{ digits})$
	2...19,99M Ω	0,01M Ω	
	20,0...599,9M Ω	0,1M Ω	

- adjustable limit
- adjustable measurement time
- automatic discharge of the capacity of the tested device after measurement
- protection against measuring live devices

Measurement of differential leakage current:

Display range	Resolution	Accuracy
0,00...3,99mA	0,01mA	$\pm(5\% \text{ m.v.} + 2 \text{ digits})$
4,0...19,9mA	0,1mA	

- adjustable limit
- adjustable measurement time

Measurement of substitute leakage current:

Display range	Resolution	Accuracy
0,00...3,99mA	0,01mA	$\pm(5\% \text{ m.v.} + 2 \text{ digits})$
4,0...19,9mA	0,1mA	

- adjustable limit
- adjustable measurement time
- open circuit voltage 25...50 V

Measurement of touch leakage current:

Display range	Resolution	Accuracy
0,00...4,999mA	0,001mA	$\pm(5\% \text{ m.v.} + 3 \text{ digits})$

- adjustable limit
- adjustable measurement time

Measurements of RCD parameters

RCD trip time test t_A

Test range according to IEC 61557: 0 ms ... to the upper limit of displayed value

RCD type	Factor	Range	Resolution	Accuracy
General	$0,5 \cdot I_{Dn}$	0...300 ms	1 ms	$\pm(2\% \text{ m.v.} + 2 \text{ digits}^1)$
	$1 \cdot I_{Dn}$			
	$2 \cdot I_{Dn}$			
	$5 \cdot I_{Dn}$			

¹⁾ - accuracy of differential leakage current $I_{Dn} = 10 \text{ mA}$ | $0,5 \cdot I_{Dn}$: $\pm 2\% \text{ m.v.} \pm 3 \text{ digits}$

Measurement of RCD disconnection current I_A for sinusoidal differential current

Test range according to IEC 61557: $(0,3...1,0)I_{Dn}$

Selected nominal current of RCD	Test range	Resolution	Test current	Basic uncertainty
10 mA	3,3...10,0 mA	0,1 mA	$0,3 \times I_{Dn} \dots 1,0 \times I_{Dn}$	$\pm 5\% I_{Dn}$
15 mA	4,5...15,0 mA			
30 mA	9,0...30,0 mA			

- it is possible to start the measurement from the positive of the negative half of forced leakage current
- test current passage time max. 3200 ms
- automatic measurement of RCD disconnection time (t_A) and disconnection current
- measurement for: $0,5I_{Dn}$, $1I_{Dn}$, $2I_{Dn}$ i $5I_{Dn}$,

Measurement of power S:

Display range	Resolution	Accuracy
0...999 VA	1 VA	$\pm(5\% \text{ m.v.} + 3 \text{ digits})$
1,00 k...3,99 kVA	0,01 kVA	

Measurement of power P:

Display range	Resolution	Accuracy
0...999 W	1 W	$\pm(5\% \text{ m.v.} + 3 \text{ digits})$
1,00 k...3,99 kW	0,01 kW	

Power factor PF:

Display range	Resolution	Accuracy
0,00...1,00	0,01	$\pm(10\% \text{ m.v.} + 5 \text{ digits})$

Measurement of current consumption:

Display range	Resolution	Accuracy
0,00...15,99 A	0,01 A	$\pm(2\% \text{ m.v.} + 3 \text{ digits})$

Voltage measurement:

Display range	Resolution	Accuracy
195,0...265,0 V	0,01 V	$\pm(2\% \text{ m.v.} + 2 \text{ digits})$



PAT-820/815/810

Innovative, intuitive touch interface:

2014/07/29 5:45:45 Admin $U_{NPE} = 2.4 \text{ V } U_{LN} = 233.1 \text{ V } f = 50.0 \text{ Hz}$

Perform test

Rpe	Isub	Power test
Riso	Ipe	RCD
IA	IEC	
Visual check	IT	Flash test

Home

2014/07/29 5:45:24 Admin $U_{NPE} = 2.5 \text{ V } U_{LN} = 232.2 \text{ V } f = 50.0 \text{ Hz}$

Sonel\Office

Objects	Devices
Office	123/14 Printer
Service	123/15 Computer

Home

2014/07/29 5:46:41 Admin

R_{ISO} - Insulation resistance

TEST IN PROGRESS $I_{ISO} = 0.00 \text{ mA}$
 $U_{ISO} = 105 \text{ V}$
 $R_{ISO MIN} > 99.9 \text{ M}\Omega$

2014/07/29 5:46:42

3 s

Test voltage U _{iso}	Test duration t	Limit	Test method
100 V	5 s	1.00 MΩ	Probe-socket

Home

2014/07/29 5:53:53 Admin

R_{PE} - PE continuity

READY!

✓ $R_{PE} = 0.23 \Omega$

2014/07/29 5:53:51

Positive test result

Test current I	Test duration t	Limit	Test method
0.2 A	35 s	0.5 Ω	Probe-socket

Home

2014/07/29 5:48:26 Admin $U_{NPE} = 1.7 \text{ V } U_{LN} = 234.6 \text{ V } f = 50.1 \text{ Hz}$

\Sonel\Office\Printer

Evidence No	123/14	2014/07/29 5:48:09 Admin SONEL PAT-820 (SN.:BJ0014)
labelName	Printer	
Producer		
Model		
Serial No		
Year of production	0	
Class		
Cycle	6	

Date of retest **2015/01/29**

Home

2014/07/29 5:48:46 Admin $U_{NPE} = 1.6 \text{ V } U_{LN} = 232.9 \text{ V } f = 50.1 \text{ Hz}$

Printer:Test history 1/2

Visual check

2014/07/29 5:47:36

- ✓ Plug
- ✓ Case
- ✓ Safety Features
- ✓ Lead
- ✓ Mechanical

R_{PE}

2014/07/29 5:47:57

✓ $R_{PE} = 0.04 \Omega$ **R_{PE}Limit = 1.0 Ω** $I_n = 0.2 \text{ A}$

R_{ISO}

2014/07/29 5:48:09

✓ $R_{ISO} > 599.9 \text{ M}\Omega$ **R_{ISO}Limit = 1.00 MΩ** $U_n = 500 \text{ V}$

Home

Conforms to the EMC requirements according to EN 61326-1:2013 and EN 61326-2-2:2013.

Electrical safety:

- measurement category
- enclosure protection rating acc. to EN 60529:

Other technical specification:

- power supply
- load current
- data transmission to PC
- dimensions: 195 V...265V, 50Hz max. 16A (230V) USB 2.0
- weight approx. 6.2 kg
- operating temperature: -10°...+50°C
- storage temperature: -20°...+70°C
- humidity 20..80%
- display TFT 7" 800x480

Standard accessories:

- power supply cord
- test lead 1,8 m; SP-4 plug, orange
- test lead with banana plug; 1.8 m; 5 kV; red (only PAT-820 - 2 pcs.)
- pin probe 5kV with banana connector - red (only PAT-820 - 2 pcs.)
- USB cable
- fuse 0314 015.VXP 15 A 250 VAC 6.3x32 mm Littlefuse 2 pcs
- instruction manual
- warranty card

WAPRZZAS1
WAPRZ1X80RKS
WAPRZ1X8REBB
WASONRE0GB2
WAPRZUSB
WAPOZB15PAT

Additional accessories:

- adapter IEC 60320 C6 Plug to IEC 60320 C13 Connector Block
- current clamp C-3 (only for PAT-820/815)
- three phase socket adapter 16A
- three phase socket adapter 16A switchable
- three phase socket adapter 32A
- three phase socket adapter 32A switchable
- adapter for industrial sockets 16A
- adapter for industrial sockets 32A
- cable - adapter Shuko / IEC (for testing extensions)
- Sonel PAT - software
- USB barcode reader
- portable USB report/bar code printer

WAADAPATIEC1
WACEGC30KR
WAADAPAT16P
WAADAPAT16PR
WAADAPAT32P
WAADAPAT32PR
WAADAPAT16F1
WAADAPAT32F1
WAADAPATIEC2
WAPROSONPAT2
WAADACK2D
WAADAD2